



MANDATORY SERVICE BULLETIN

DATE: August 11, 2011 Service Bulletin No. 2011-1
SUBJECT: DigiWx Inspection Cycle Change
MODELS AFFECTED: Belfort DigiWx AV100
TIME OF COMPLIANCE: At next periodic inspection

The FAA has approved a change in the inspection cycle for the Belfort DigiWx AWOS AV. The inspection cycle has been changed to every 120 days and the pressure sensor calibration check is now every 12 months. The previous inspection cycle was every 90 days including the check of the pressure sensors. You can download the revised DigiWx maintenance manual and FAA authorization letter at the Belfort web site www.belfortinstrument.com under DigiWx.

It is recommended that this inspection program be incorporated at your earliest convenience.

Belfort Field Service Team
800 937-2353
service@belfortinstrument.com